

SPECIFICATIONS

SHIELDING REQUIREMENTS:

- BACK WALL 1/2" THK LEAD
- SIDE WALLS 3/8" THK LEAD
- ROOF 1/4" THK LEAD

CONSTRUCTION: MODULAR

STRUCTURE: SELF-SUPPORTED

- ROOF RATED: 40 LBS LIVE WEIGHT/SQ FT
- 100 LBS DEAD WEIGHT/SQ FT

OVERALL DIMS:

- 3.5M WIDE X 7.5M LONG X 5.8M HIGH [11.5' X 24.6' X 19.0']
- CLEAR HEIGHT INSIDE: 5.5M [18']
- DOOR OPENING: 2.4M HIGH X 1.0M WIDE [8' X 3.3']

SLIDING DOORS:

- ONE MANUAL, FIXED BY LATCH
- ONE PNEUMATIC WITH MANUAL OVERRIDE

SIDE PANEL: 1.0M WIDE X 5.8M HIGH [3.3' X 19.0']

- SANDWICH CONSTRUCTION: STEEL 3mm [1/8" (11Ga.)] - LEAD 10mm[3/8"] - STEEL 6mm[1/4"]
- STRUCTURE: 3" X 3" X 1/4" STEEL ANGLE, ALL AROUND
- WEIGHT:2567LB

BACK PANEL: 1.0M WIDE X 5.8M HIGH [3.3' X 19.0']

- SANDWICH CONSTRUCTION: STEEL 3mm [1/8" (11Ga.)] - LEAD 13mm[1/2"] - STEEL 6mm[1/4"]
- STRUCTURE: 3" X 3" X 1/4" STEEL ANGLE, ALL AROUND
- WEIGHT: 3030LB

STANDARD ROOF PANEL: 1.0M WIDE X 3.5M LONG [3.3' X 11.5']

- SANDWICH CONSTRUCTION: STEEL 3mm [1/8" (11Ga.)] - LEAD 6mm[1/4"] - STEEL 6mm[1/4"]
- STRUCTURE: MC 6 X 3" X 16.3 STEEL CHANNEL, ALL AROUND
- WEIGHT:1714LB

STANDARD DOOR PANEL: 1.0M WIDE X 2.5M HIGH [3.3' X 8']

- SANDWICH CONSTRUCTION: STEEL 3mm [1/8" (11Ga.)] - LEAD 10mm[3/8"] - STEEL 6mm[1/4"]
- WEIGHT:1165LB

LABYRINTH PANELS: UTILITIES, CABLES, VENTILATION


- UTILITIES: CHILLED WATER 5° - 0.8°C [41° - 1.5°F]
- COMPRESSED AIR 90psi
- ELECTRICAL POWER 30kVA 120/208V

WORK STATION CRANE WITH MANUAL HOIST:

- SPAN 3.0 [10.0'] X RUNWAY LENGTH 7.2M [23.5'] X CLEAR HEIGHT 5.5M [18.0']
- CAPACITY: 1000kg [1 TON]

VENTILATION: 2 FANS - 545 CFM, 8.2 SONES @ FREE AIR

ILLUMINATION: FLUORESCENT - OFFICE RATED: 100 FOOT CANDLE.



NOTES:

1- THE HUTCH MANUFACTURER SHALL BE RESPONSIBLE FOR THE STRUCTURAL INTEGRITY OF THE HUTCH, AND SHALL CERTIFY THAT THE STRUCTURE WAS REVIEWED BY A LICENSED STRUCTURAL ENGINEER.

STRUCTURES AND THEIR ELEMENTS SHALL BE BUILT FOLLOWING DOE 6430 AS REFERRED TO IN TECHNICAL SPECS. DESIGN LOAD SHALL INCLUDE LOADS THAT CAN OCCUR DURING CONSTRUCTION. DEAD LOADS SHALL INCLUDE THE USE OF THE INTERIOR UNISTRUT FOR SHELVES ONE FOOT WIDE SUPPORTING LOADS OF 200 LBS.

2- LEAD THICKNESS IS SPECIFIED BY ARGONNE. THE MANUFACTURER SHALL BE RESPONSIBLE FOR PROVIDING A RADIATION LEAK PROOF HUTCH.

LEAD LINED PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE 'C' TO COMPLY WITH FEDERAL SPECIFICATION QQ-L-201. HUTCH SHALL MEET ALL APPLICABLE RADIATION GUIDE LINES: REFERRED TO IN TECHNICAL SPECS.

REV	DESCRIPTION	DATE	BY	CHECKED BY	DATE
001	TO COMPLETE ASSY DRAWINGS	03/09	CSM		
002	REVISED	03/09	CSM		

NO.	NAME	DATE	NO.	NAME	DATE
1	A1873701		1	ARGONNE NATIONAL LABORATORY	
2	CORCIERA	02/09	2	D. SHU	03/04
3	AL MACRANDER	03/04	3	T. KUZAY	04/04
4	SANCHEZCOR	03/04	4	D. MILLS	04/04
5	T. SANCHEZ	03/04	5		

DESCRIPTION	DATE	BY	CHECKED BY
PERSPECTIVE & SPECS		N/A	
		03/09	

PROJECT NO.	P41050903-610000-01
-------------	---------------------