



NOTE:  
LEAD PANEL SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION QQ-289. HUTCH SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, AND N43.2 AND DOE ORDER 5480.11

ITEM	QTY/PART	DESCRIPTION	SYMBOL	DATE	BY	DATE	BY	MATERIAL / SPEC	REV
<b>PARTS LIST</b>									
<b>A2051401</b>									
<b>ARGONNE NATIONAL LABORATORY</b>									
<b>ADVANCED PHOTON SOURCE</b>									
H1 MBES ROOF # 5 1/4" x 3.00M x 1.00M									
REVISIONS									
01	CHANGE TO SIZE D AND MAKE ROOM FOR 2 UNISTR.	TSAC	JC	6/13/95					