







- NOTES:
1. LEAD PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION OQL201. CABINET SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANS N432 AND DOE ORDER 5480.11.
 2. PAINT EXPOSED LEAD SURFACES PER SPEC. (EXCEPT AS NOTED)
 3. DIMENSIONS IN BRACKETS () ARE MILLIMETERS AND ARE FOR REFERENCE ONLY.
 4. SOME HIDDEN LINES MAY BE OMITTED FOR CLARITY.



A2291200		ARGONNE NATIONAL LABORATORY	
DATE	02/28/05	PROJECT	U6 BM SHIELDED CABINET L4 & P8 SHUTTER & SLITS ASSEMBLY
DESIGNED BY	SANCHEZ/CORCORAN	CHECKED BY	SANCHEZ/CORCORAN
SCALE	1/8"=1	SECTION	E
P4105090606-300000-00			