



SIDING PREPARATION TYPICAL DETAIL
 SCALE= 1:2



- NOTES:
- LEAD PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION-QLQ201. CABINET SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANS N43.2 AND DOE ORDER 5480.11.
 - FINISH - PAINT EXTERIOR SURFACES, EXCEPT AS NOTED, WITH PRATT & LAMBERT #2127 (SNO-FLAKE WHITE, SEMI-GLOSS, EPOXY BASE) OR EQUIVALENT. SURFACE PREPARATION PER A.N.L. SPECIFICATION #4105-000 24.
 - DIMENSIONS IN BRACKETS () ARE MILLIMETERS AND ARE FOR REFERENCE ONLY.
 - SOME HIDDEN LINES MAY BE OMITTED FOR CLARITY.

ITEM	QUANTITY	SYMPART NUMBER	MANUFACTURE OR DESCRIPTION	MATERIAL / SPEC	REV
PARTS LIST					
PARTS LISTED IN THIS DRAWING ALL DIMENSIONS ARE IN INCHES UNLESS OTHERWISE SPECIFIED FINISHES 1. 2011 102 2. 2011 250 3. 2011 175			PARTS NOT LISTED IN THIS DRAWING 1. 2011 102 2. 2011 250 3. 2011 175		
PROJECT NUMBER A2225300		DATE 5/10/08		DRAWN BY corcuera	
CHECKED BY SANCHEZ/CORCUERA		DATE 5/10/08		PROJECT NAME U6 BM SHIELDED CABINET DIAGNOSTICS ROOF PANEL	
SEE NOTES				SCALE 1:4	SHEET NUMBER D 14
DRAWING NUMBER P4105090606-700004-00					

DATE	BY	DESCRIPTION