



**SIDING PREPARATION TYPICAL DETAIL**  
SCALE= 1:2



- NOTES:
- LEAD PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION-QLQ201. CABINET SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANS N43.2 AND DOE ORDER 5480.11.
  - FINISH - PAINT EXTERIOR SURFACES, EXCEPT AS NOTED, WITH PRATT & LAMBERT #2127 (SNO-FLAKE WHITE, SEMI-GLOSS, EPOXY BASE) OR EQUIVALENT. SURFACE PREPARATION PER A.N.L. SPECIFICATION #4105-000 24.
  - DIMENSIONS IN BRACKETS ( ) ARE MILLIMETERS AND ARE FOR REFERENCE ONLY.
  - SOME HIDDEN LINES MAY BE OMITTED FOR CLARITY.

ITEM	QUANTITY	SYMPART NUMBER	NOMENCLATURE OR DESCRIPTION	MATERIAL / SPEC	REV
<b>PARTS LIST</b>					
PARTS LISTED IN THIS DRAWING ALL DIMENSIONS ARE IN INCHES UNLESS OTHERWISE SPECIFIED			THIS DRAWING IS THE PROPERTY OF <b>ARGONNE NATIONAL LABORATORY</b>		
DRAWN BY <b>M. PATAKY</b>		DATE <b>5/1/95</b>	CHECKED BY <b>SP</b>		DATE <b>5/1/95</b>
PROJECT NO. <b>5111</b>		PROJECT NAME <b>ADVANCED PHOTON SOURCE</b>		U6 ID SHIELDED CABINET DIAGNOSTICS ASSEMBLY ROOF PANEL	
SEE NOTES					
DRAWN BY: M. PATAKY CHECKED BY: SP				SCALE: 1/4 SHEET: 14	PART NUMBER: <b>D P4105090606-900004-00</b>

REV	DATE	DESCRIPTION